

Calibration Certification



Cert #:	27787-1
Cal Date:	30/11/2023
Recall Date:	29/11/2024
Interval:	12 months

Customer:	VESTAS Commande N°4509003521
Contact:	Julien FIOLE
Address:	VESTAS AREA WHS FRANCE MAIN TOOLS WAREHOUSE RUE DES AZALÉS 10800 SAINT THIBAUT

Phone:	+33 6 37 247 212
Fax:	
P.O.:	4509003521

Test Instrument:	Torque Transducer	Mfgr:	AWS Series, Inc.
Model #:	41014-7.5K-HY	Procedure:	TTC-98-03
S/N:	27787-1	ID/Lot #:	M277871
Capacity:	7500	State Accuracy:	+/- 0.5 % of indicated reading, top
Units:	lb ft	Range:	90 % of FS ☐ Extended Range

	Nominal	As Found	Out	As Left	Tolerance (+/-)
Clockwise	750.00			748.90	3.7500
	1500.0			1499.5	7.5000
	3000.0			2999.9	15.000
	4500.0			4500.8	22.500
	6000.0			6001.1	30.000
	7500.0			7502.3	37.500
CounterClockwise	-750.00			-751.20	3.7500
	-1500.0			-1501.5	7.5000
	-3000.0			-3001.2	15.000
	-4500.0			-4503.5	22.500
	-6000.0			-6005.2	30.000
	-7500.0			-7506.1	37.500

CALIBRATION EQUIPMENT USED

MODEL NO.	MANUFACTURER	ITEM	S/N	Recall
TTCA-10K	AWS, Inc.	Arm	119350	11/23/2024
VARIOUS	Troemmer	Weights	W-300	02/03/2024

683/289870-17, 683/285541-14

WE CERTIFY THAT THIS TEST INSTRUMENT HAS BEEN COMPARED TO STANDARDS TRACEABLE TO N.I.S.T AND/OR NPL AND HAS BEEN CALIBRATED TO THE STATED ACCURACY. CALIBRATION DATE AND RECALL DATE FOR NEW EQUIPMENT IS BASED ON ORIGINAL DATE OF MFG. ACTUAL RECALL DATE IS 6 TO 24 MONTHS FROM RECEIPT OF NEW EQUIPMENT AS APPLICABLE.

Remarks: New Unit

Supervisor:

David HUSTACHE

Technician:

David Hinkel

69210 FLEURIEUX SUR L'ARBRESLE

Tel.: 04 78 33 39 19

E-mail: sn@hytorc-ce.com

N° SIRET: 808 881 189 00027

TVA: FR 11 808 881 189

Cert Date: 30/11/2023

Ship Date:



ADVANCED WITNESS SERIES, HYTORC HUSTACH. 179 RUE DE MONTÉPY 69210 FLEURIEUX SUR L'ARBRESLE FRANCE.
+33 4 78 33 39 19 contact@hytorc-ce.com hytorc.hustach.com