

GAUGE CALIBRATION CERTIFICATE

Date Issued: 26/02/2025

Certificate Number: 004003

Gauge Serial Number: 103889-13-25

Gauge Scale: 2000 bar 25000 psi

Test Gauge Serial Number: 924103

Pressure Gauge [bar]	Test Gauge [bar]	Correction [bar]	Pressure Gauge [psi]	Test Gauge [psi]	Correction [psi]	% Error FSD
0	0	0	0	0	0	0.0%
200	210	-10	2900	3045	-145	-0.5%
400	410	-10	5800	5945	-145	-0.5%
600	610	-10	8700	8845	-145	-0.5%
800	810	-10	11600	11745	-145	-0.5%
1000	1010	-10	14500	14645	-145	-0.5%
1200	1210	-10	17400	17545	-145	-0.5%
1400	1410	-10	20300	20445	-145	-0.5%
1500	1520	-20	21750	22040	-290	-1.0%

Accuracy of test gauge $\pm 0.1\%$

This gauge was found to be accurate to within $\pm 1\%$ of full scale deflection.
Pressure gauge was tapped before each reading.

Authorised signature:



HYTORC
Hustach

HYTORC-Hustach - 179 Rue de Montépy ZA de Montépy 69210 FLEURIEUX-SUR-L'ARBRESLE
+33 4 78 33 39 19 - contact@hytorc-ce.com - hytorc-hustach.com - e-torc.com - hytorc-services.com